



RPF. No. 4903 /Estt.(BBMC&H)/23.

Dt. 17-09-23

**TENDER NOTICE FOR SUPPLY OF TWO NOS. OF DENTAL RVG MACHINE FOR XRAY
OF DENTAL DEPARTMENT OF BHIMA BHOI MEDICAL COLLEGE & HOSPITAL,
BALANGIR-767001.
PROJECT SANCTIONED UNDER MCL CSR FUND.**

Bid Reference No. - 4903, Dt. 17-09-23

Date of Issue of Tender:-

Dt.17.09.2023

Last date & Time of submission of Tender:-

Dt.09.10.2023 & 04.00 P.M.

Date & Time of Pre Bid Meeting & Presentation:-

Dt.26.09.2023 & 11.00 A.M.

Date & Time of opening of Technical Bid:-

Dt.10.10.2023 & 11.00 A.M.

Date & Time of opening of Financial Bid:-

Dt.10.10.2023 & 01.00 P.M.

M. Park
17/09/23



SALE OF TENDER / BID DOCUMENT

The Bidders may download the Tender Documents directly from the WEBSITE available at <https://bbmchbalangir.nic.in>. The Tender cost fee of Rs.1000/- (inclusive of Postage Charges, Non-refundable) by way of separate Demand Draft drawn in favour **Superintendent, Bhima Bhoi Medical College & Hospital, Balangir** payable at **Balangir** should be enclosed along-with the Bid. The Bidders should specifically superscribe, “**Tender for Supply of 02 nos. of Dental RVG Machine**” on the top left corner of the outer envelope containing the Bid. The Tender cost fee and the EMD amount should be submitted separately in separate demand drafts. In case of any bid amendment and clarification, responsibility lies with the bidders to collect the same from the website or the office notice board. The BBMC&H, Balangir shall have no responsibility for any delay / omission on part of the bidder.

Price of bidding document: **Rs.1000.00** (inclusive Postage Charges)
(Rupees One Thousand) only
(Non-refundable)

The tender paper will be rejected if the bidder changes any clause or Annexure of the bid document downloaded from the website.

M.Pandey
17/09/23

17/09/23



**TERMS AND CONDITIONS FOR SUPPLY OF DENTAL RVG MACHINE FOR
DENTAL X-RAY**

- 1.1 Sealed tenders will be received by Dated **09.10.2023** upto **4:00 PM** by the **Superintendent, Bhima Bhoi Medical College & Hospital, Balangir**, Odisha for the purchase of DENTAL XRAY (RVG). Any tender received after the due date & time will be rejected / returned unopened to the sender. The tenders will be received through Regd. Post / Courier services / Speed Post only. The tender should mention valid phone number and e mail –id over the envelop.
- 1.2 The Sealed tenders submitted by the tenderers will be opened by the selection committee of BBMC&H, Balangir in his office at **11.00 A.M.** on **10.10.2023**. The tenderer or their duly authorized representatives are allowed to be present during the opening of the tenders, if they so like.

A. Specification

RVG For Dental X-Ray

Technical Specification:

Size 1: Technology Super : SMOS, SCINTILLATOR Optical Fiber

Pixel size : 19

True (measured) resolution : 16 ip/mm

Active area : 22.2 x 29.6 mm

Ovarall dimension: 27,6 x 37.7 mm

Sensor plate thickness: 7.3 mm

Gray scale : 4096 grey level (12 bits)

Input voltage (from USB interface): 5V

Proper training is provided at the time of installation with user's manual

Presentation:

- Presentation of the above item will be given by each bidder before the committee on dtd.**26.09.2023** at **11.00 A.M** regarding specification & Compatibility of the above item.

B. Documents to be submitted along with Bids:

1. Technical Bid shall be furnished as per the Specification given and should be submitted in hard copy.
2. Price Bid, should include price of all taxes & transportation charges.
3. The tenderer must submit Earnest Money Deposit (EMD) of Rs. **7000/- (Rupees Seven Thousand)** only in favor of Superintendent, Bhima Bhoi Medical College & Hospital, Balangir in shape of Bank Draft from a Scheduled Bank payable at Balangir. The EMD of the unsuccessful bidders will be returned back after finalization of Bid. In case of bidder has authentic MSME registration Certificate, bidder may avail exemption of EMD by submitting the valid MSME Certificate.

M.Pandey
13/10/23



C. Notes:

If the firm fails to supply the Item in time or the item supplied are of Substandard quality or violates the tender terms and conditions the EMD amount will be forfeited and the firm will be blacklisted for two years.

D. Tender Terms and Conditions :

- a) The selected firms will have to supply within 15 days from issue of purchase order.
- b) The item should arrive at the Superintendent, Bhima Bhoi Medical College & Hospital, Balangir in good condition.
- c) The rate will remain valid for one year from the date of approval.
- d) In the event of the date being declared as a holiday for purchaser's Office, the due date of submission of bids & opening of bids will be the following working date & time.
- f) The authority reserves the right to accept / reject any part thereof or all the bids & without assigning any reason thereof.

Selection Procedure:

The bidders meeting all the specification Criteria and Compatibility with the Dental X-Ray Machine available in the dept. will be considered for the technical qualification. The bidders qualifying the technical qualification will be consider for financial evaluation. The technically qualified bidders quoting the lowest bid price will be selected as L1 bidder for awarding of the contract.

E. Issue of purchase order:

- a. The purchase order will be placed to the concerned firms after the approval of the purchase committee.

F. Payment:

100% payment will be made after receipt of stock entry and successful test run of the equipment and submission of Installation Certificate.

Sd/-

Superintendent,

Bhima Bhoi Medical College & Hospital, Balangir

M/10/12



FINANCIAL PROPOSAL

RFP NO: 4903, Date: 17-09-23

Sl. No.	Name of Item and Specification	Required Quantity	Unit price inclusive of all taxes, transportation and installation charges.
1	Dental RVG Machine for X-Ray Technical Specification: Size 1: Technology Super: SMOS, SCINTILLATOR Optical Fiber Pixel size: 19 True (measured) resolution: 16 ip/mm Active area: 22.2 x 29.6 mm Overall dimension: 27,6 x 37.7 mm Sensor plate thickness: 7.3 mm Gray scale: 4096 grey level (12 bits) Input voltage (from USB interface): 5V Proper training is provided at the time of installation with user's manual	02 nos.	

Signature of Bidder

M.P. Panigrahi
17/09/23