



**OFFICE OF THE SUPERINTENDENT BHIMA BHOI  
MEDICAL COLLEGE & HOSPITAL**

At/PO/Dist – Balangir, PIN – 767001 (Odisha)

E-mail: [gmchospitalbalangir1@gmail.com](mailto:gmchospitalbalangir1@gmail.com)



\*\*\*

Tender Ref. No. IX-F-36 1693 /Estt.(BBMC&H).

Dt. 27.02.2024

**TENDER DOCUMENT FOR SUPPLY OF INSTRUMENT & EQUIPMENTS TO BHIMA BHOI  
MEDICAL COLLEGE & HOSPITAL, BALANGIR FOR THE YEAR 2024-25.**

1. DATE OF AVAILABILITY OF BID DOCUMENT IN WEBSITE: **From Date :28.02.2024**
2. PRE BID CONFERENCE : **Date:11.03.2024 at 11.30AM**
3. LAST DATE OF AVAILABILITY OF BID DOCUMENTS  
(In website) : **Date.20.03.2024**
4. LAST DATE & TIME OF RECEIPT OF BID DOCUMENTS : **Dt.20.03.2024 Up to 1.00PM**
5. DATE & TIME OF OPENING OF COVER-A (Technical Bid): **Dt.21.03.2024 11.00AM**
6. DATE OF OPENING OF COVER-B (Price Bid) : **will be intimated Latter**
7. PLACE OF OPENING OF BID DOCUMENTS AND  
ADDRESS FOR COMMUNICATION  
AND RECEIPT OF BID DOCUMENTS : **O/o Superintendent, Bhima Bhoi  
Medical Collage& Hospital,  
District-Balangir-767001.**

*M. Pande*  
27-02-24  
Superintendent

Bhima Bhoi Medical College & Hospital  
Balangir - 767001



### TENDER / BID DOCUMENT

The Bidders have to download the Tender Documents directly from the Website available at college web portal <https://bbmchbalangir.nic.in>, Balangir district [www.balangir.nic.in](http://www.balangir.nic.in) & DMET web portal [www.dmetodisha.gov.in](http://www.dmetodisha.gov.in). The Tender cost fee of Rs.2,000/- (Two Thousand) Non-refundable by way of separate Demand Draft drawn in favour of Superintendent, Bhima Bhoi Medical College & Hospital, Balangir payable at Balangir should be enclosed along-with the Technical Bid (Cover-A). The Bidders should specifically super scribe, **"Supply of Instrument & Equipments for Bhima Bhoi Medical College & Hospital, Balangir for the Year 2024-25"** on the top left corner of the outer envelope containing Technical Bid (Cover-A) and Price Bid (Cover-B) separately. The Technical Bid (Cover-A) & Financial Bid (Cover-B) should be kept within the outer Cover. The outer sealed envelope must be sent through **Speed Post/Registered Post/By Courier facility (not by hand)** only. The tender document will not be received/entertained after due date & time fixed for receipt in the advertisement and will be rejected summarily. The Tender cost fee and the EMD amount should be submitted separately in shape of demand drafts in the technical bid. In case of any bid amendment and clarification, responsibility lies with the bidders to collect the same from the website or Sub store, Bhima Bhoi Medical College & Hospital, Balangir before last date of submission of tender document and the authority shall have no responsibility for any delay/omission on part of the bidder. The validity of the tender is one year from the date of finalization of the tender process.

*Tender Paper Cost:*

Rs. 2,000/-  
(Non-refundable)

\*\*\*\*\* The tender paper will be rejected if the bidder changes any clause or Annexure of the bid document downloaded from the website \*\*\*\*\*

*M. Pandey*  
*27-02-24*

Superintendent  
Bhima Bhoi Medical College & Hospital  
Balangir - 767001



**IMPORTANT INSTRUCTIONS TO BE NOTED CAREFULLY BY THE TENDERERS**

1.	Mode of Procurement	Through Open Tender
2.	Purchaser	Superintendent, Bhima Bhoi Medical College & Hospital, Balangir.
3.	Consignee	Addl. Medical Superintendent (Store), Bhima Bhoi Medical College & Hospital, Balangir.
4.	Delivery Period	Within 15 days from issue of Purchase Order.
5.	Mode of Delivery	By Air / Road / Rail
6.	EMD	<b>Rs. 50,000/- (Rupees Fifty Thousand)</b> only per tender. The Earnest Money Deposit should be in the shape of demand Draft/Banker's Cheque only in favour of Superintendent, Bhima Bhoi Medical College & Hospital, Balangir drawn in any Nationalized / Scheduled Bank payable at BALANGIR. EMD exemption is not permitted except to local MSME/SSI units registered in Odisha only. The MSME certificate( Category- Manufacturer Only) of EMD exemption to be produced.
7.	Performance Security	<p>The supplier has to submit Performance Security of <b>Rs. 50,000/- (Rupees Fifty Thousand)</b> in shape of DD/Bank Guarantee of the order Value within 15days of issue of supply order. The performance Security is to be refunded after completion of the validity period of the Tender or 1 year from the supply of such items whichever is earlier without any interest. The performance security will be forfeited in the following cases: -</p> <ul style="list-style-type: none"><li>i. Any violation of the terms &amp; conditions of Tender.</li><li>ii. If the items supplied is not of standard quality.</li><li>iii. Blacklisting of the Agency/Supplier.</li></ul>

M. Pandey  
27.02.24

Superintendent

Bhima Bhoi Medical College & Hospital  
Balangir - 767001



**TERMS AND CONDITIONS FOR SUPPLY OF INSTRUMENT & EQUIPMENTS FOR BHIMA  
BHOI MEDICAL COLLEGE & HOSPITAL, BALANGIR FOR THE YEAR 2024-25.**

1. Tender Fee of Rs.2,000/- (Rupees Two thousand only) (Non-refundable) should be submitted in shape of Demand Draft issued from a Nationalized Bank in favour of Superintendent, Bhima Bhoi Medical College & Hospital, Balangir.
2. EMD of **Rs. 50,000/- (Rupees Fifty thousand only)** should be submitted in shape of Demand Draft issued from a Nationalized Bank in favour of Superintendent, Bhima Bhoi Medical College & Hospital, Balangir. However, the EMD of the unsuccessful bidders will be returned after completion/finalization of the tender process and for successful bidders the EMD amount will be kept as Performance security. The performance security amount will be returned without interest after completion of the tender tenure with satisfactory performance. In case of selected Local MSME manufacturer having valid MSME certificate (Category-Manufacturer), who have claimed exemption of EMD, have to deposit performance security of 50,000/- (Rupees Fifty thousand only) in shape of DD/Bank Guarantee in favour of Superintendent, Bhima Bhoi Medical College & Hospital, Balangir. For Local MSME (Category-Manufacturer) selected manufacturer, the performance security amount will be returned after completion of tender tenure.
3. The bidder should quote the List of items as mentioned in **ANEXURE-VII** in the format attached in **ANNEXURE-1 (A)**. It is mandatory to produce the Catalogue for all the items in the list.
4. Copy of valid GST Registration Certificate should be attached with the Tender Documents.
5. Copy of Valid PAN Card should be attached with the tender documents.
6. The Bidder should submit valid Manufacturing authorization certificate in original in the name of the Bidder for this Tender only for each item of Instrument & Equipment from Manufacturer with letter no. mentioned in letter head of the manufacturer (**ANNEXURE-II**).
7. The Bidder should submit the valid manufacturing licence of medical devices.
8. Copy of valid **ISO Certificate** of the manufacturer should be attached with the tender documents.
9. Copy of Valid CE (**Conformite Europeenne**) Certificate of the manufacturer should be attached with the tender documents.
10. Authorized Distributors on behalf of the Manufacturers are eligible to participate in the tender provided:
  - i. They submit Manufacturer's authorization and Power of Attorney to transact business on behalf of the Manufacturer. The authorized Distributor may raise bill, if specially authorized by the Manufacturer.
  - ii. The authorized distributor will submit all the documents in support of eligibility of the Manufacturer as mentioned above.
11. Bidder (Manufacturer/Distributor) should have proof of supply of at least 05 nos. of the quoted Instrument & Equipment's to any Govt. Organization/Corporate Hospital/PSU Hospital/UN Agency during last 3 financial years (2020-21,2021-22,2022-23). The Experience/ End User certificate in support of supply must be submitted along with the bid document.
12. Proof of Average Annual Turnover of Bidder (Manufacturers/Distributor) of **Rs.2 Crore or more** during the financial year (**2020-21,2021-22,2022-23**) taken together from a registered Chartered Accountant shall be submitted.

*M Pande*  
27.02.24  
Superintendent  
Bhima Bhoi Medical College & Hospital  
Balangir - 767001



13. The copy of the Audited financial statement by Registered Chartered Accountant including **Balance sheet, Profit & Loss account** showing for the last three financial years (2020-21,2021-22,2022-23) shall be submitted.
14. **CE (Conformite Europeenne) Certification logo** on the body of the instrument must be present.
15. Manufacturer Name/Logo on the body of the instrument must be present.
16. If any instruments found to be defective/not of standard Quality, the same shall be replaced by the supplier at his own cost within a maximum duration of 30 days and take back the unused/defective available stock from the Sub-store. In case the supplier fails to replace the stocks within the stipulated time, the **performance security** will be forfeited. No further purchase order will be placed to the firm/ supplier for the item (s) and the firm/supplier will be blacklisted/debarred from participating in any tender (for that item) floated in future for three years.
17. For the, Items mentioned in the list attached in **Annexure-VII (S/L No- 239,240,241,242,243,276,277,278,279,280,281,339)** Samples of such items will be verified by the Technical Committee and according to the suggestion, the samples which don't qualify in sample verification will not be considered during price bid Comparison. The Proforma for submission of samples is attached in **ANNEXURE-1(B)**.
18. The date of sample verification will be intimated after opening of the technical bid document. The Technical committee constituted by the Tendering authority will verify the samples.
19. The bidder will have to come along with the samples on the date of sample verification and the sample will be returned to them after completion of the sample verification on the same date. Non submission of the samples will lead to technical disqualification for these quoted items and will not be taken into consideration in the comparative statement.
20. The Technically qualified bidder quoting the lowest rate for each item will be selected as L1 bidder for award of contract and next qualified bidder quoting next higher rate will become L2, L3..... accordingly.
21. If the approved lowest eligible supplier fails to supply items within 30 days of receiving supply order, the same shall be procured from the L2 suppliers at L1 rate, if they agree to supply at L1 approved rate to meet the emergency requirement.
22. The goods should be supplied to the Freight on Road (FOR) Destination and no extra charge for freight should be claimed.
23. The tender format (**Price Bid**) in the prescribed format (as per **ANNEXURE-III**), must be submitted inside sealed price Bid cover. The rate should be quoted for each item in figures only and not in words. Only two decimal points (paise) will be taken into consideration ignoring the rest digits.
24. The rate should be mentioned excluding GST. The GST amount & Total amount for each item should be mentioned in a separate column in the same sheet.

M. Panda  
27.02.24

5 | Page

Superintendent  
Bhima Bhoi Medical College & Hospital  
Balangir - 767001



25. Tenders will be evaluated as per the specification mentioned in the list of items attached in List of Items **Annexure-VII** and price bid will be opened only for the bidders who qualify in the Technical Bid. The cost of the unit item excluding GST will be evaluated/Compared.
26. Copy of the Declaration/undertaking as per format **ANNEXURE-V** should be attached with the tender documents.
27. The authority is not responsible for any type of postal delay. The tender document will not be received/entertained after due date & time fixed for receipt in the advertisement and will be rejected summarily.
28. The tenderer should submit/furnish a certificate in the tender to the effect that, “the **price quoted by them is not more than the open market price or also under GeM Rate Contract where such rate exists**”. However, in circumstances when the price decreases during the contract period, the approved supplier, should intimate the same to the purchasing authority **(ANNEXURE-VI)**.
29. The tender documents should be clearly written/types without any correction, interpolations and overwriting. Each page of the tender should bear the dated signature & stamp of the tenderer.
30. All copies of the tender document should be self-attested and stamped. If any information or documents furnished by the tenderer are found to be incorrect or misleading at any stage, the tender will not be considered for evaluated and rejected.
31. All the documents shall be serially numbered and self-attested by the bidder.
32. The approved rate will be valid for **one year** from the date of approval or till finalization of the next tender whichever is later. It may be extended for further period as per the decision of the committee/Tendering authority.
33. The quantum of procurement will be made on requirement basis.
34. The authority reserves the right to cancel/reject all the bids or any part thereof without assigning any reason thereof.

*M. Pande*  
*27-02-24*  
Superintendent  
Bhima Bhoi Medical College & Hospital  
Balangir - 767001



**ANNEXURE-I**

Sl No	Particulars	Submitted		Documents (Page No.)
		Yes	No	
1	Details of Manufacturing Unit/contract person/ Lisoning agent/servicing center in Odisha			
2	Tender paper cost (Non Refundable)			
3	Earnest Money Deposit			
4	Affidavit duly certified by Notary regarding Nonblack listing of the Agency/Firm. (As per ANNEXURE-IV)			
5	List of Quoted Item(in given format ANNEXURE-1A), (Referred to ANNEXURE-VII) for total List of 348 Items			
6	Manufacturer's Authorization Certificate in original for distributor/Wholesaler ANNEXURE-II.			
7	Copy of Valid manufacturing License for Medical Devices of Manufacturer.			
8	Certified Copy of Valid ISO Certificate of Manufacturer			
9	Certified Copy of Valid CE certificate of the Manufacturer			
10	Photo copy of Valid PAN Card			
11	Photo copy of valid GST Registration			
12	Photocopy of GST Return for the last Quarter shall be submitted (Nov-23, Dec-23 & Jan-24).			
13	Proof of Average Annual Turnover (Rs. 2 Crore or more )for last three financial years (2020-21, 2021-22,2022-23) duly signed by the Regd. Chartered Accountant.			
14	Audited Financial Statement by Regd. C.A. of last 3 F.Ys (2020-21, 2021-22,2022-23) which includes Balance sheet, shall be submitted			
15	Performance report/End User Certificate during the last 3 F. Ys (2020-21,2021-22,2022-23) regarding supply of 05 Nos of quoted items to any Govt. Organization/Corporate Hospital/PSU Hospital /UN Agency) Purchase order in support shall be submitted.			
16	Copy of the Declaration (ANNEXURE-V)			
17	Copy of original Tender and schedules, duly signed by the Bidder			
18	Certificate/undertaking of Price not more than the market price or als under GeM Rate Contract where such rate exists. (ANNEXURE-VI)			

**Important: 1. Bidders are instructed to submit their Bids as per given Format (List of Instrument & Equipment's) including serial number against mentioned items otherwise their Bids will be rejected**

**2. Particulars for all of the 18 parameter is mandatory and shortfall/Deficiency of any one or more in the Checklist will lead to technical disqualification of the bidder.**

\*\*\*

M. Panda  
27-02-24

7 | Page

Superintendent  
Bhima Bhoi Medical College & Hospital  
Balangir - 767001

**ANNEXURE – I (A)**  
**LIST OF ITEMS QUOTED**

(To be attached with technical bid)

Category – A (Instrument & Equipment)

SI No	Item Sl. No. as per tender	Name of the Item	Specification	Name of the manufacturer	Product certification (CE / ISI /Manufacturer Name or Logo) copy attached)	ISO certified company (if yes provide a copy)	Catalogue/Broucher
1					CE – ISI –	Pg.-	Pg-
2					CE – ISI –	Pg.-	Pg-



**ANNEXURE-I (B)**

**PROFORMA FOR SUBMISSION OF SAMPLES**

SI No	Item Sl. No. as per tender	Name of the Item	Specification (CE/ISI etc)	Name of the manufacturer / Brand	Quantity
1					
2					

(THE SCHEDULE DATE OF SAMPLE VERIFICATION WILL BE INTIMATED ON THE DATE OF OPENING OF TECHNICAL BID/ LATER ONBY EMAIL ONLY)

**N.B:**

1. The bidder shall come along with the Samples on the date of sample verification.
2. Each sample shall be properly wrapped, labeled with manufacturer, size, brand etc. for smooth verification.

M. Pandey  
27-02-24  
Superintendent  
Bhima Bhoi Medical College & Hospital  
Balangir - 767001



**MANUFACTURER'S AUTHORISATION FORMAT**

(in original in case bidder is authorized by manufacturer)

Letter No. \_\_\_\_\_

Date. \_\_\_\_\_

To

The Superintendent  
Bhima Bhoi Medical College & Hospital, Balangir

**Sub: LETTER OF AUTHORISATION.**

Ref: Tender Reference No.1693 Dtd.27.02.2024.

Madam/Sir,

We \_\_\_\_\_ who are  
established and reputed manufacturer's of \_\_\_\_\_  
(name and description of items offered) having factories  
at \_\_\_\_\_

\_\_\_\_\_ (Address of Factory) do hereby  
authorize M/s \_\_\_\_\_

(Name and address of Distributor / Agent) to submit a bid and sign the contract with you  
against the above referred tender.

We also extend our full quality assurance for the items quoted by M/s \_\_\_\_\_

\_\_\_\_\_ as per the terms and conditions in your tender  
under reference above.

Yours faithfully,

Full Name of the designated person

(Signature with seal)

Contact Number:

Email ID:

**Note:** This letter of authority should be on the letter head of the manufacturer and should be signed by a person competent and having the power of attorney to bind the manufacturer. It should be included in the bid submitted by the tenderer if the tenderer is not the manufacturer.



PRICE SCHEDULE (PRICE BID)

ANEXURE-III

Sl. No.	Item Sl. No. as per tender	Name of the item	Make / model / Specification	Name of the manufacturer	Each rate (Rs.)	GST%	Total unit cost including tax

*MPand*  
*27-02-24*  
 Superintendent  
 Bhima Bhoi Medical College & Hospital  
 Balangir - 767001



**ANNEXURE – IV**

**DECLARATION FORM**

I/We \_\_\_\_\_ having  
My/Our \_\_\_\_\_  
office at \_\_\_\_\_ do declare that  
I/We have carefully read all the terms & conditions of the tender of the \_\_\_\_\_, Orissa  
for the supply of Instrument & Equipment. I will abide with all the terms & conditions set forth in the tender  
paper reference no. \_\_\_\_\_.

**I/We do hereby declare I/We have not been de-recognized / black listed by any State Govt. / Union  
Territory / Govt. of India / Govt. organization / Govt. Health Institutions for supply of Not of Standard  
Quality(NSQ) items / part-supply / non-supply.**

*I/We agree that the Tender Inviting Authority can forfeit the Earnest Money Deposit and or performance  
Security Deposit and blacklist me/us for a period of 3 years if, any information furnished by us proved to be false  
at the time of inspection / verification and not complying with the Tender terms & conditions.*

I / We further declare that I / We possess valid manufacturing license / authorized distributor license bearing  
No. \_\_\_\_\_ Valid upto \_\_\_\_\_ I / We  
\_\_\_\_\_ do hereby declare that  
I / we will supply the \_\_\_\_\_ as per the terms,  
conditions & specifications of the tender document. I / we further declare that my / our EMD / performance security  
will be forfeited if I / we fail to supply any item after getting order from the purchaser.

Signature of the bidder :

Date :

Name & Address of the Firm: Affidavit before Executive Magistrate / Notary Public.



ANNEXURE-V

DECLARATION / UNDERTAKING

I/We \_\_\_\_\_ having My/our  
\_\_\_\_\_ office at  
\_\_\_\_\_ do hereby declare that I/We will supply  
the ordered items after becoming lowest responsive bidder as per tender terms, conditions,  
specification [ Bid Reference No: **1693 Dtd.27.02.2024** and conditions as laid down in the purchase  
order.

I/We declare that, the Tender Inviting Authority **can blacklist me/us for a period of 3 years** if we  
withdraw bid after opening of price bid and / or approval of rate contract or unable to supply ordered items  
at approved rate within the stipulated time period.

Signature of the bidder :

Name:

Mobile No.:

Date :

Name & Address of the Firm: Affidavit before **Executive Magistrate / Notary Public.**

*MPande*  
*27-02-24*

Superintendent

Bhima Bhoi Medical College & Hospital  
Balangir - 767001



ANNEXURE-VI

DECLARATION / UNDERTAKING

I/We \_\_\_\_\_ having My/our  
\_\_\_\_\_ office at  
\_\_\_\_\_ do hereby declare and undertake that the  
price quoted by us for the items Vide Tender Ref. No.1693 Dtd.27.02.2024 is not more than the open  
market Price, MRP and also under GeM Rate contract where such rates exists, However when the  
price decreases during the contract period, we will ethically intimates the same to the Purchasing  
authority.

Signature of the bidder :

Name:

Mobile No.:

Date :

M. Panda  
27-02-24  
Superintendent  
Bhima Bhoi Medical College & Hospital  
Balangir - 767001



**ANNEXURE-VII (LIST OF ITEMS)**

SL. NO	NAME OF THE ITEMS	SPECIFICATION	Rate Quoted
1	Abdominal Retractor Stainless steel	Stainless Steel CE Marking	Each
2	Adsons Forcep Blunt 3 Inch Stainless steel	Stainless Steel CE Marking	Each
3	Adsons Forcep Blunt 4 Inch Stainless steel	Stainless Steel CE Marking	Each
4	Adsons Forcep Blunt 5 Inch Stainless steel	Stainless Steel CE Marking	Each
5	Adson's forceps 6 inches	Stainless Steel CE Marking	Each
6	Adsons Forcep Tooth 3 Inch Stainless steel	Stainless Steel CE Marking	Each
7	Adsons Forcep Tooth 4 Inch Stainless steel	Stainless Steel CE Marking	Each
8	Adsons Forcep Tooth 5 Inch Stainless steel	Stainless Steel CE Marking	Each
9	Adult Weight Machine		Each
10	Digital Adult Weight Machine		Each
11	Allis Forceps (4-5 teeth) size-3 Inch Stainless steel	Stainless Steel CE Marking	Each
12	Allis Forceps (4-5 teeth) size-4 Inch Stainless steel	Stainless Steel CE Marking	Each
13	Allis Forceps (4-5 teeth) size-5 Inch Stainless steel	Stainless Steel CE Marking	Each
14	Allis Forceps (4-5 teeth) size-6 Inch Stainless steel	Stainless Steel CE Marking	Each
15	Allis Forceps (4-5 teeth) size-8 Inch Stainless steel	Stainless Steel CE Marking	Each
16	Artery Forcep Straight 3 Inch Stainless steel	Stainless Steel CE Marking	Each
17	Artery Forcep Straight 4 Inch Stainless steel	Stainless Steel CE Marking	Each
18	Artery Forcep Straight 5 Inch Stainless steel	Stainless Steel CE Marking	Each
19	Artery Forcep Straight 6 Inch Stainless steel	Stainless Steel CE Marking	Each
20	Artery Forcep Straight 8 Inch Stainless steel	Stainless Steel CE Marking	Each
21	Artery Forceps Curved size 3 Inch Stainless steel	Stainless Steel CE Marking	Each
22	Artery Forceps Curved size 4 Inch Stainless steel	Stainless Steel CE Marking	Each
23	Artery Forceps Curved size 5 Inch Stainless steel	Stainless Steel CE Marking	Each
24	Artery Forceps Curved size 6 Inch Stainless steel	Stainless Steel CE Marking	Each
25	Artery Forceps Curved size 8 Inch Stainless steel	Stainless Steel CE Marking	Each
26	B.P. Handle size-3	Stainless Steel CE Marking	Each
27	B.P. Handle size-4 (Scalpel Handle 24 Size)	Stainless Steel CE Marking	Each
28	Babcock Forcep 4 Inch Stainless steel	Stainless Steel CE Marking	Each
29	Babcock Forcep 5 Inch Stainless steel	Stainless Steel CE Marking	Each
30	Babcock Forcep 6 Inch Stainless steel	Stainless Steel CE Marking	Each
31	Babcock Forcep 8 Inch Stainless steel	Stainless Steel CE Marking	Each
32	Back Rest	Stainless Steel CE Marking	Each
33	Bone Cutter	Stainless Steel CE Marking	Each
34	Bowel Small Size	Stainless Steel CE Marking	Each
35	Bowel Big Size	Stainless Steel CE Marking	Each
36	Bull Dog Clamp (Normal)	Stainless Steel CE Marking	Each
37	Bull Dog Clamp (T-Shape)	Stainless Steel CE Marking	Each

*M. Pande*  
27-02-24

Superintendent

Bhima Bhoi Medical College & Hospital  
Balangir - 767001



38	Cheatle Forcep Stand	Stainless Steel CE Marking	Each
39	Cheatle Forceps Size-4 Inch Stainless steel	Stainless Steel CE Marking	Each
40	Cheatle Forceps Size-5 Inch Stainless steel	Stainless Steel CE Marking	Each
41	Cheatle Forceps Size-6 Inch Stainless steel	Stainless Steel CE Marking	Each
42	Cheatle Forceps Size-8 Inch Stainless steel	Stainless Steel CE Marking	Each
43	Cheatle Forceps Size-10 Inch Stainless steel	Stainless Steel CE Marking	Each
44	Cockers Straight Size-5 Inch Stainless steel	Stainless Steel CE Marking	Each
45	Cockers Straight Size-6 Inch Stainless steel	Stainless Steel CE Marking	Each
46	Cockers Straight Size-8 Inch Stainless steel	Stainless Steel CE Marking	Each
47	Cord Cutting Scissor Stainless steel	Stainless Steel CE Marking	Each
48	Diverse Retractor Medium Stainless steel	Stainless Steel CE Marking	Each
49	Doyen Abdominal Retractor	Length-23.5 Cm, Width-7.5cm, Thickness-37x31mm	<b>Each</b>
50	Diverse Retractor Small Stainless steel	Stainless Steel CE Marking	Each
51	Epistomy Scissor Stainless steel	Stainless Steel CE Marking	Each
52	Fine Curved Scissor Size-3 Inch Stainless steel	Stainless Steel CE Marking	Each
53	Fine Curved Scissor Size-4 Inch Stainless steel	Stainless Steel CE Marking	Each
54	Fine Curved Scissor Size-5 Inch Stainless steel	Stainless Steel CE Marking	Each
55	Fine Curved Scissor Size-6 Inch Stainless steel	Stainless Steel CE Marking	Each
56	Fine Scissor Tissue Cutting Stainless steel	Stainless Steel CE Marking	Each
57	Galli Pot 3 Inch Stainless steel	Stainless Steel CE Marking	Each
58	Galli Pot 4 Inch Stainless steel	Stainless Steel CE Marking	Each
59	Galli Pot 5 Inch Stainless steel	Stainless Steel CE Marking	Each
60	Galli Pot 6 Inch Stainless steel	Stainless Steel CE Marking	Each
61	Green Armytage Forcep Size-4 Inch Stainless steel	Stainless Steel CE Marking	Each
62	Green Armytage Forcep Size-5 Inch Stainless steel	Stainless Steel CE Marking	Each
63	Green Armytage Forcep Size-6 Inch Stainless steel	Stainless Steel CE Marking	Each
64	Green Armytage Forcep Size-8 Inch Stainless steel	Stainless Steel CE Marking	Each
65	Hynis Instrument Set	Stainless Steel CE Marking	Each Set
66	Kidney Tray Small 6 Inch Stainless steel	Stainless Steel CE Marking	Each
67	Kidney Tray Size- 8 Inch Stainless steel	Stainless Steel CE Marking	Each
68	Kidney Tray Size-10 Inch Stainless steel	Stainless Steel CE Marking	Each
69	Kidney Tray Size-12 Inch Stainless steel	Stainless Steel CE Marking	Each
70	Kochers Forceps (Curved) 6 Inch Stainless steel	Stainless Steel CE Marking	Each
71	Kochers Forceps (Curved) 8 Inch Stainless steel	Stainless Steel CE Marking	Each
72	Kochers Forceps (Straight) 6 Inch Stainless steel	Stainless Steel CE Marking	Each
73	Kochers Forceps (Straight) 8 Inch Stainless steel	Stainless Steel CE Marking	Each
74	Lona Curved Fine Scissor	Stainless Steel CE Marking	Each
75	Lona Curved Thin Artery Forcep	Stainless Steel CE Marking	Each
76	Mayo Tissue Cutting Scissor Size-4 Inch Stainless steel	Stainless Steel CE Marking	Each
77	Mayo Tissue Cutting Scissor Size-5 Inch Stainless steel	Stainless Steel CE Marking	Each

M Pande  
27.02.24



78	Mayo Tissue Cutting Scissor Size-6 Inch Stainless steel	Stainless Steel CE Marking	Each
79	Mayo Tissue Cutting Scissor Size-7.5 Inch Stainless steel	Stainless Steel CE Marking	Each
80	Mayo Tissue Cutting Scissor Size-8 Inch Stainless steel	Stainless Steel CE Marking	Each
81	Mayo Tissue Cutting Scissor Size-10 Inch Stainless steel	Stainless Steel CE Marking	Each
82	Mayo Tissue (Curved) Scissor Size-12 Inch Stainless steel	Stainless Steel CE Marking	Each
83	Metallic Tongue Depressor Small		Each
84	Metallic Tongue Depressor Medium		Each
85	Metallic Tongue Depressor Large		Each
86	Mosquito Artery Forcep Curved Size-3 Inch Stainless steel	Stainless Steel CE Marking	Each
87	Mosquito Artery Forcep Curved Size-3.5 Inch Stainless steel	Stainless Steel CE Marking	Each
88	Mosquito Artery Forcep Curved Size-4 Inch Stainless steel	Stainless Steel CE Marking	Each
89	Mosquito Artery Forcep Curved Size-5 Inch Stainless steel	Stainless Steel CE Marking	Each
90	Mosquito Artery Forcep Curved Size-6 Inch Stainless steel	Stainless Steel CE Marking	Each
91	Mosquito Artery Forcep Straight Size-3 Inch Stainless steel	Stainless Steel CE Marking	Each
92	Mosquito Artery Forcep Straight Size-4 Inch Stainless steel	Stainless Steel CE Marking	Each
93	Mosquito Artery Forcep Straight Size-5 Inch Stainless steel	Stainless Steel CE Marking	Each
94	Mosquito Artery Forcep Straight Size-6 Inch Stainless steel	Stainless Steel CE Marking	Each
95	Needle Holder Straight 2500 teeth/ Sq. Inch Stainless steel-10 Inch Stainless steels	Stainless Steel CE Marking	Each
96	Needle Holder Straight 2500 teeth/ Sq. Inch Stainless steel-3 Inch Stainless steels	Stainless Steel CE Marking	Each
97	Needle Holder Straight 2500 teeth/ Sq. Inch Stainless steel-4 Inch Stainless steels	Stainless Steel CE Marking	Each
98	Needle Holder Straight 2500 teeth/ Sq. Inch Stainless steel-5 Inch Stainless steels	Stainless Steel CE Marking	Each
99	Needle Holder Straight -6 Inch Stainless steels	Stainless Steel CE Marking	Each
100	Needle Holder Straight -8 Inch Stainless steels	Stainless Steel CE Marking	Each
101	Needle holder heavy 6 inches	Stainless Steel CE Marking	Each
102	Needle holder heavy 8 inches	Stainless Steel CE Marking	Each
103	Plain Forceps Size-4 Inch Stainless steel	Stainless Steel CE Marking	Each
104	Plain Forceps Size-5 Inch Stainless steel	Stainless Steel CE Marking	Each
105	Plain Forceps Size-6 Inch Stainless steel	Stainless Steel CE Marking	Each
106	Plain Forceps Size-8 Inch Stainless steel	Stainless Steel CE Marking	Each
107	Plain Scissor 4 Inch Stainless steel	Stainless Steel CE Marking	Each
108	Plain Scissor 5 Inch Stainless steel	Stainless Steel CE Marking	Each
109	Plain Scissor 6 Inch Stainless steel	Stainless Steel CE Marking	Each
110	Plain Scissor 8 Inch Stainless steel	Stainless Steel CE Marking	Each
111	Proctoscope Clamps (Adult)	Stainless Steel CE Marking	Each
112	Proctoscope Clamps (Small For Child)	Stainless Steel CE Marking	Each



113	Rectangular Steel Tray 10 x 8 Inch Stainless steel	Stainless Steel CE Marking	Each
114	Rectangular Steel Tray 12x10 Inch Stainless steel	Stainless Steel CE Marking	Each
115	Right Angle Retractor Small 4 Inch Stainless steel	Stainless Steel CE Marking	Each
116	Right Angle Retractor Small 5 Inch Stainless steel	Stainless Steel CE Marking	Each
117	Right Angle Retractor Small 6 Inch Stainless steel	Stainless Steel CE Marking	Each
118	Right Angle Retractor 8 Inch Stainless steel	Stainless Steel CE Marking	Each
119	Scissor Straight Size-6 Inch Stainless steel	Stainless Steel CE Marking	Each
120	Scissor Straight Size-8 Inch Stainless steel	Stainless Steel CE Marking	Each
121	Scissor Long Curved 4 Inch Stainless steel	Stainless Steel CE Marking	Each
122	Scissor Long Curved 5 Inch Stainless steel	Stainless Steel CE Marking	Each
123	Scissor Long Curved 6 Inch Stainless steel	Stainless Steel CE Marking	Each
124	Scissor Long Curved 8 Inch Stainless steel	Stainless Steel CE Marking	Each
125	Scissor Long Straight 4 Inch Stainless steel	Stainless Steel CE Marking	Each
126	Scissor Long Straight 5 Inch Stainless steel	Stainless Steel CE Marking	Each
127	Scissor Long Straight 6 Inch Stainless steel	Stainless Steel CE Marking	Each
128	Scissor Long Straight 8 Inch Stainless steel	Stainless Steel CE Marking	Each
129	Scissor Short Curved 4 Inch Stainless steel	Stainless Steel CE Marking	Each
130	Scissor Short Curved 5 Inch Stainless steel	Stainless Steel CE Marking	Each
131	Scissor Short Curved 6 Inch Stainless steel	Stainless Steel CE Marking	Each
132	Scissor Short Curved 8 Inch Stainless steel	Stainless Steel CE Marking	Each
133	Single Blade Speculum	Stainless Steel CE Marking	Each
134	Double Blade Speculum	Stainless Steel CE Marking	Each
135	Sinus Forcep 3 Inch Stainless steel	Stainless Steel CE Marking	Each
136	Sinus Forcep 4 Inch Stainless steel	Stainless Steel CE Marking	Each
137	Sinus Forcep 5 Inch Stainless steel	Stainless Steel CE Marking	Each
138	Sinus Forcep 6 Inch Stainless steel	Stainless Steel CE Marking	Each
139	Sims Speculum	Stainless Steel CE Marking	Each
140	Cosco Speculum	Stainless Steel CE Marking	Each
141	Speculum Small Size	Stainless Steel CE Marking	Each
142	Sponge Holding Forcep 4 Inch Stainless steel	Stainless Steel CE Marking	Each
143	Sponge Holding Forcep 5 Inch Stainless steel	Stainless Steel CE Marking	Each
144	Sponge Holding Forcep 6 Inch Stainless steel	Stainless Steel CE Marking	Each
145	Sponge Holding Forcep 8 Inch Stainless steel	Stainless Steel CE Marking	Each
146	Sponge Holding Forcep 10 Inch Stainless steel	Stainless Steel CE Marking	Each
147	Sponge Holding Forcep Straight 10 Inch Stainless steel	Stainless Steel CE Marking	Each

M Panda  
27.02.24

Superintendent

Bhima Bhoi Medical College & Hospital  
Balangir - 767001



148	Stethoscope (Adult & Paediatric)	<p>1-Stethoscope of standard size, chromium plated metal binaural, V rubber tube in one piece and rotating piper fitting for both flip functions.2-Double sided adult &amp; paediatric stethoscope.3-Extra-soft, replaceable and pivot able ear-tips for perfect sealing at the ear canal.4-Designed with precision chest-piece made of stainless steel/ chromed brass.5-Good quality diaphragm of maximum -Ø 45mm.6-High quality membrane for precise acoustics with non-chill rims for improved adaptation on the skin and for excellent sound transmission.7-Length should be 27" to 29" with colour - black.8-The Y-tube should be made of Latex-free treated rubber.</p> <p>Easy to dismantle, and therefore to clean and disinfect.</p> <p><b>PHYSICAL CHARACTERISTICS:</b>1-Dimensions (metric)- Diaphragm size as specified.2-Weight (lbs, kg)-Weight: 110-150 gm.3-Mobility, portability-Yes.</p> <p><b>ACCESSORIES, SPARE PARTS,</b></p> <p><b>CONSUMABLES:</b>1. Accessories (mandatory, standard, optional)- 1 x spare set of earpiece, 1 x spare diaphragm</p>	Each
149	Sterilized Drum Size-11x9 Inch Stainless steel	Stainless Steel CE Marking	Each
150	Sterilized Drum Size-9x9 Inch Stainless steel	Stainless Steel CE Marking	Each
151	Stitch Cutting Scissor	Stainless Steel CE Marking	Each
152	Suture Cutting Scissor	Stainless Steel CE Marking	Each
153	Tailor's Scissor 12 Inch Stainless steel (Big)	Stainless Steel CE Marking	Each
154	Disposable Tetra	Stainless Steel CE Marking	Each
155	Tissue Cutting Scissor	Stainless Steel CE Marking	Each
156	Tooth Forceps 3 Inch Stainless steel	Stainless Steel CE Marking	Each
157	Tooth Forceps 4 Inch Stainless steel	Stainless Steel CE Marking	Each
158	Tooth Forceps 5 Inch Stainless steel	Stainless Steel CE Marking	Each
159	Tooth Forceps 6 Inch Stainless steel	Stainless Steel CE Marking	Each
160	Tooth Forceps 8 Inch Stainless steel	Stainless Steel CE Marking	Each
161	Tooth Forceps (Large) 10 Inch Stainless steel	Stainless Steel CE Marking	Each
162	Torch Light		Each
163	Towel Clip	Stainless Steel CE Marking	Each
164	Baby Tray 18x12	Stainless Steel CE Marking	Each
165	Instrument Tray 15x12	Stainless Steel CE Marking	Each
166	Tray Small 10X8 Inch Stainless steel	Stainless Steel CE Marking	Each
167	Tray Small 8X6 Inch Stainless steel	Stainless Steel CE Marking	Each
168	Urethral Dilatation (Metallic)		Each
169	Urethral Dilatation (Teflon)		Each
170	Dilator Sets	Stainless Steel CE Marking	Each

M Pande  
27-02-24

Superintendent

Bhima Bhoi Medical College & Hospital  
Balangir - 767001



171	Uterine Sound	Stainless Steel CE Marking	Each
172	Uterine Curate 8 Inch Stainless steel	Stainless Steel CE Marking	Each
173	Uterine Curate 10 Inch Stainless steel	Stainless Steel CE Marking	Each
174	Uterine Clamp 8 Inch Stainless steel	Stainless Steel CE Marking	Each
175	Uterine Clamp 10 Inch Stainless steel	Stainless Steel CE Marking	Each
176	Venesection Set	Stainless Steel CE Marking	Each SET
177	Digital Haemoglobinometer (BIOSENSE)		Each
178	Myoma Screw		Each
179	Tubal Recanalisation Set		Each
180	Ventouse Cup		Each
181	Outlet Forceps		Each
182	INSTRUMENT STERILIZER (ELECTRICAL)	1-Seamless shell & lever operated Lid fitted with full proof mechanism control excessive steam escape and restricts condensate within the shell.2- Incredible synchrosed manoeuvrability of lid, due to statistically perforated tray for flushing and entry of water.3- Configuration design and material as per IS:5022.4- Material 304 Grade SS Sizes-Large-(510x200x150mm), Medium-(430x200x150mm), Small-(300x150x125mm)	Each
183	INSTRUMENT STERILIZER (ELECTRICAL)	Made of Stainless steel 202 Grade Without tray lifting arrangement Size-Large-(510x200x150mm), Medium-(430x200x150mm), Small-(300x150x125mm)	
184	Dressing Trolley	1-Overall approx. size- 760mmLx500mmWx900mmH, 2-Approximate shelf dimension 750mmLx500mmW, 3- Tubular CRC frame mounted on four castors of minimum 100mm dia and should be made of stainless steel.4-Two S.S. of 304 grade Shelves with protective railing on three sides.5-Should have provision for holding bowl and bucket.6- Should be provided with heavy duty SS bowl & SS bucket of thickness 20G.	Each
<b>GENERAL SURGERY</b>			
185	Linear GI Stapler With cartridge	Blue-100mm/3.8mm, Green-80mm/4.8mm	Each
186	Skin Hook (Adult)	Stainless steel 6 Inches	Each Pair
187	Intestinal Clamp (non-crushing)	Straight	Each
188	Intestinal Clamp (non-crushing)	Fehland	Each
189	Skin Grafting Blade	Name-Esemann standard	Each
190	watsons skin grafting handle	Stainless steel 310mm	Each

M. Panda  
27.02.24



191	Trucut automatic Biopsy Gun	Size-16G, 10cm length	Each
192	Trucut Biopsy Needle (manual)		Each
193	Stevenson Tenotomy Scissor	Stainless steel size-straight sharp-4.25 Inch, Curved sharp-4.25 inch super cut	Each
194	Czerny Retractor		Each
195	SPC Trocar	Size-16,18	Each
196	Torumo guide wire	Size-150cm-0.035 inch	Each
197	Urology Teflon serial dilator		Each
198	Tweezer Forcep (fine)	Name-Adson Microforcep Size-6 Inches,4 Inches	Each
199	Twin intestinal clamp (non-crushing)		Each
200	Cooley's Bull dog Forcep	Stainless steel Size-90mm/3.50 inches	Each
201	Facial Nerve Dissector		Each
202	Weitlaner Retractor (2x3 Prongs Blunt)		Each
203	Weitlaner Retractor (2x3 Prongs Sharp)		Each
204	Weitlaner Retractor (3x4 Prongs Blunt)		Each
205	Weitlaner Retractor (3x4 Prongs Blunt)		Each
206	ICD Drain Clamper		Each
207	satinskys vascular clamp ( large)		Each
208	satinskys vascular clamp ( medium)		Each
209	satinskys vascular clamp ( small)		Each
210	Drain Milker		Each
211	Bull dog curve vascular clamp medium and large size		Each
212	Abdominal Mopping Tetra		Each
<b>LAPROSCOPIC INSTRUMENTS</b>			
213	Laparoscopic Maryland Bipolar Cautery		Each
214	Safety Trocar automatic (5mm)		Each
215	Safety Trocar automatic (10mm)		Each
216	Vessel Sealer		Each
217	Haemostatic Forcep		Each
218	V Lock Suture		Each
219	Grasper Plane		Each
220	Grasper Tooth		Each
221	10mm Port		Each Pair
222	5mm Port		Each Pair
223	Veriss Needle		Each
224	Port Closure Instrument		Each
225	Laparoscopic Scissor	Stony/Striker Shaft length-33cm, Diameter-10mm, Shaft length-45cm diameter-10mm	Each Pair
226	Clip Applicator	Stainless steel, Clamp, Diam and dimension-10mmx330um, Size-standard 300/400 clamp	Each



227	Laprosopic knot pusher stainless steel closed end 5mmx35cm		Each
228	harmonic vessel sealer		
229	gigglesaw		
230	Laprosopic Marylane Bipolar Vessel sealer & cutter	Stainless steel 330x5mm	Each

### ANESTHESIOLOGY

231	Magilis Forcep Paediatric Size-3 Inch Stainless steel	Stainless Steel CE Marking	Each
232	Magilis Forcep Paediatric Size-4 Inch Stainless steel	Stainless Steel CE Marking	Each
233	Magilis Forcep Adult Size-6 Inch Stainless steel	Stainless Steel CE Marking	Each
234	Magilis Forcep Adult Size-8 Inch Stainless steel	Stainless Steel CE Marking	Each
235	Laryngoscope (Anterior Commissure Laryngoscope)	With Fibre Optic carrier (Small, Medium, Large)	Each
236	Laryngoscope With 3 Number Straight Blade	Size-2-1,3-1,4-1	Each
237	Laryngoscope With 4 number Curved Blade	Size-2-1,3-1,4-1,5-1	Each
238	Macoy Laryngoscope With 3 Blade Curved	Size-2-1,3-1,4-1	Each

239	Portable Video Laryngoscope With Laryngoscope Blades	Receives images from distal tip of video adaptor and displays image on integrated full-color anti-glare screen	Each
		Portable- Battery operated	
		Rigid video Laryngoscope: Integrated reusable video 3,5 inch multi-function 180 degree rotating display TFT LCD screen (Thin Film Transistor Liquid Crystal Display) with On/Off switch, Battery indicator and video-out port	
		Reusable video adaptors (contains an in-built LED light source and CMOS sensor waterproof camera housed in distal tip)	
		Disposable concave laryngoscope blades that slides over the video adaptor and connect to the reusable display unit. Distal window of blade has anti-fog coating to prevent moisture condensation affecting image in display unit. Two types of blades : (i) Channeled blade-has a guiding channel for fast delivery of ET tube into the vocal cords; (ii) Standard blade-has no guiding channel. Blade sizes: 1[ intabts<4yrs], 2 [1-10 yrs], 3[>5 yrs] & 4 [adults]	
		Channel to guide placement of endotracheal tube	
		Video-out-cable to connect to an external monitor (for teaching purposes & recording)	
		Power provided by 3 AAA batteries housed in battery compartment	
		Provision for video recording/ image storing	



		Instant playback of images	
		To examine and visualize the structures of upper airway and help in placement of endotracheal tube in the trachea.	
240	Peripheral NERVE Stimulator/TOF Monitor	Should be FDA & CE certified	Each
		Should provide 0.2 mA to 5.0 mA Pulse amplitude	
		Pulse width should be 0.1 m sec	
		Should have 1 Hz to 2 Hz Pulse frequency. The frequency should be easily switchable. Default frequency should be of 1 Hz	
		User should be able to toggle the Frequency between 1Hz and 2Hz by holding the concerned key for 2 seconds	
		Default frequency of 1 Hz should be Audible only during Self-testing and afterwards will be triggered only when the circuit is closed	
		Impedance should be of 0-5 K Ohms	
		Should have 3 digit 7 segment large LED display unit	
		Should provide 0.01 mA - 0.1 mA resolution	
		Should be compatible with 9V alkaline BF battery/ batteries available locally	
		Should have Red LED flashing light to indicate low battery	
		Should have Constant current Stimulator-monophasic and audio-visual indicators for pulse delivery	
		Should have yellow/green LED flashing light with beep sound for delivery of each Pulse when the circuit is closed	
		Should have Digital LED display of frequency with beep tone to display impulse current from 0.20 to 5.0 mA	
		Should have Pulse LED (yellow / green) to indicate readiness of device	
		Stimulus duration should be of 0.05 ms – 0.10 ms – 0.30 ms – 0.50 ms – 1.00 ms +/- 1%	
		Allowable load impedance should be 0 k Ohms – 12 k Ohms	
		Current measuring accuracy should be +/- 0.02 mA	
		Impedance measuring range should be 1 K Ohms – 90 K Ohms for target stimulation current >0.5 mA	
		Impedance measuring accuracy should be +/- 10% or +/-20% for target stimulation	
		Should be operable at 0-50°C and 90% relative humidity without any condensation	
		Should consume not more than 6mA-8mA power	



		Maximum stimulation current should be 5mA (0-12 K Ohms)	
		Pulse duration should be changeable between 0.1, 0.3 and 1.0 ms to target motor/sensory nerve components	
		Should have audible Disconnection and Current thresholds alarms	
		Should deliver Single twitch stimulation at 0.1Hz (1 Stimulus every 10 seconds), Train of Four(TOF) at 2Hz (2 stimulus per second) and Tetanic stimulation at 50Hz (50 stimulus per second)	

		Positive and negative electrodes should be colour coded	
		Should be user friendly, have low maintenance, transportable	
		User manual and troubleshooting guide should be provided	
		Fifty (50) number of Insulated Needles should be provided with the instrument	
		Use should be demonstrated to user by technical person at time of installation	
		Should provide the necessary circuit attachments and leads as required for operating the machine.	
		Should come with 1 year guarantee.	

### ENT INSTRUMENT

241	LED Head Light Source With Fiber optic Cable	(3.5mm,4.5mm), 7.5 ft 100/120 watt., Lifespan at least 60,00 Hrs.	Each
242	LED Head Light With Rechargeable Spare Battery	(10 Watt, rechargeable spare battery)	Each
243	Portable Endoscopy unit with TFT Monitor	(TFT Monitor 16 inch Screen, Full HD 1080P Camera, Inbuilt 100 watt LED Light source with fibre optic cable	Each
244	Boyle Davis Mouth gag with Tongue piece	(Metallic with Bipod stand ( rings) with 5 Blades	Each
245	Cutting Mastoid Burr	( 3.5,4,4.5,5,5.5,6) mm, Tungsten	Each
246	Diamond Burrs	(3,4mm)	Each
247	Right Angled Forcep Metallic	Metallic	Each
248	Doyen Mouth Gag (100mm Child)		Each
249	Doyen Mouth Gag (140mm Adult)		Each
250	Metallic Tracheostomy Tube Fullers		Each
251	Metallic Tracheostomy Tube Jackson		Each
252	Insulated Alligator Forcep Nasal		Each
253	Insulated Suction Cannula (Rigid & Flexible (8 Inch Stainless steel)		Each
254	Nasal Suction Cannula No.1 & No.5		Each



255	Esophagoscope (RIGID) Adult		Each
256	Esophagoscope (RIGID) Paediatric		Each
257	Hypopharyngoscope (RIGID) Adult		Each
258	Hypopharyngoscope (RIGID) Paediatric		Each
259	Long Metallic Suction Cannula		Each
260	Esophageal Foreign Body Removal Forcep		Each
261	Fiber Optic Cable for Cold Light Source (Long Type)		Each

### DENTAL INSTRUMENT

262	Dental Diagnostic Set	With Handle No-5 :Mouth Mirror (Front Surface Mirror), Tweezers, Straight Probe, Curved Probe)	Each SET
263	Muco-Periosteal Elevator	Size-Small	Each
264	Muco-Periosteal Elevator	Size-Big	Each
265	Bone Files (Dental)		Each
266	Dental Adult Extraction Forcep Set		Each SET
267	Dental Elevator Set	Copland Elevator, Winter Cryer Elevator, Warwick James Elevator) Spear Points	Each SET
268	Luxator Tips Set (Set of 09 Nos)	3MM	Each SET
269	Tissue Retractor	Size-small	Each
270	Diamond Burrs		Each
271	surgical bur for straight hand piece		Each
272	Dental operating loupe		Each
273	Bone Rongeur	45° head 2mm TIPS	Each
274	Bone Rongeur	90° head 2mm TIPS	Each
275	<b>Dental RF Electro-Cautery Machine</b> (Radio Frequency cautery machine generates high frequency alternating AC output specially designed for surgical requirements.)	Ac Input power :225W, RF Output Power :200W, Rated output frequency: 4.2Mcycle, Intensity control: 6 setting, Unit wattage: 225W, Output waveform: alternating HF Sine wav, Hand Switch control circuit, Foot control circuit, Bipolar parameters circuit, Earth plate circuit, Shock proof ABS unit body) 2 Year Replacement Warranty)	Each

### OPHTHALMOLOGY (EYE)

276	<b>EVISCERATION SET Contains :</b>		EACH SET
	Self retaining eye speculum	Overall length-88mm, blade width-13mm,	
	Tooth Forceps	Stainless steel, 10x3x3cm, grit type-fine	
	Conjunctival Scissors	10.5cm long, blunt tips, 11mm blade curved	
	BP Blade Holder	3 & 4 inch, stainless steel 304G	
	Evisceration spatula	141mm length, round stainless steel handle	



	Needle holder	Right handle, box joint, ratchet-yes, 6, 8 inch length	
	Small straight scissors	4x4x1cm	
	<b>ENUCLEATION SET Contains :</b>		
	Self retaining eye speculum	Overall length-88mm, blade width-13mm,	
	Tooth Forceps	Stainless steel, 10x3x3cm, grit type-fine	
	Conjunctival Scissors	10.5cm long, blunt tips, 11mm blade curved	
277	Muscle hook	2mm bulbous tip with 6mm flat portion, flat serrated handle with polished finish 136mm length	EACH SET
	Optic nerve guide	21mm wide curved spoon with 5mm by 10mm notch, flat smooth handle with polished finish, 147mm length	
	Long curved scissors	curved, 8 inch length, stainless steel	
	Needle holder	Ring handle, box joint, ratchet-yes, 6, 8 inch length	
	<b>ECTROPION &amp; ENTROPION SET Contains :</b>		
278	Lid clamp (right & left)	61mm length	EACH SET
	BP Blade Holder	3 & 4 inch, stainless steel 304G	
	Small straight scissors	4x4x1cm	
	Needle holder	Ring handle, box joint, ratchet-yes, 6, 8 inch length	
	<b>CHALAZION SET Contains :</b>		
279	Chalazion clamp	3.5 inch and feature 20mm, diameter of forceps head	EACH SET
	Chalazion scoop	150mm length, flat handle, cup size-1, 2, 3, 4, 5mm	
	BP handle holder	3 & 4 inch, stainless steel 304G	
	<b>DCR SET Contains :</b>		
280	BP blade holder	3 & 4 inch, stainless steel 304G	EACH SET
	Catspaw retractor	8x7x12cm	
	small straight scissors	4x4x1cm	
	Bone gouge	stainless steel, 190mm length, 26mm tip, straight	
	Bone trephine	stainless steel, 8mm with central drill	
	Bone punch	stainless steel, straight and curved	
	Lacrimal sac guard	11x8mm, stainless steel, 4 prongs blunt	
	Needle holder	Right handle, box joint, ratchet-yes, 6, 8 inch length	
	Tooth forceps	stainless steel, 10x3x3cm, grit type-fine	
	<b>CATARACT SET Contains :</b>		
281	Universal Wire Speculum	30cm, length: 10mm blade size, blade-type-wirw, stainless steel	EACH SET
	Conjunctival Scissors	10.5cm long, blunt tips, 11mm blade curved	
	Corneal Scissors	Blunt tip, 11mm blades, 106mm length, stainless steel	



Superior Rectus holding Forceps	11cm length, fixation forceps, stainless steel
Small Artery Forceps	straight, stainless steel 5 inches
Disposable Crescent	plastic, light weight, 2mm size
Disposable Keratome	2.75mm, single bevel, plastic
Disposable Extensor	plastic, 3.2mm
Capsulorrhexis Forceps	Sharp pointed serrated interlocking tips, 110mm length, flat handle, 45 degree angled tip from shaft
Vannas Scissors	9cm long, 3mm tips, 6mm blades
Bimanual Irrigation Aspiration	stainless steel, 23g, curved thin wall shaft, 0.3mm
Cannula	aspiration, 10 inches tubing and luer-lock adapter, 15mm length
IOL Holder	107mm length, flat handle, 40 degree angled shafts
Iris Repositor	Angles, solid handle, 13cm length, 1mm working end size
IOL dialer	Angled shaft with 0.5mm tip and 0.18mm diameter, 115mm length
Needle Holder	Right handle, box joint, ratchet-yes, 6, 8 inch length

### ORTHOPEDIC INSTRUMENTS

285	Bone Hook, sharp, small, L 8 inches	Stainless Steel CE marking	Each
286	Bone Hook, large, L 10 inches	Stainless Steel CE marking	Each
287	Langenbeck Retractor, 8mm wide, L 220mm	Stainless Steel CE marking	Each
288	Langenbeck Retractor, 18mm wide, L 235mm	Stainless Steel CE marking	Each
289	Langenbeck Retractor, 24mm wide, depth 60mm, L 270mm	Stainless Steel CE marking	Each
290	Langen Back right angle Retractor 28mmx10mm	Stainless Steel CE Marking	Each
291	Langen Back right angle Retractor 40mmx11mm	Stainless Steel CE Marking	Each
292	Langen Back right angle Retractor 55mmx14mm	Stainless Steel CE Marking	Each
293	Periosteal Elevator, curved shaft, 14mm wide, L 200mm	Stainless Steel CE marking	Each
291	Periosteal Elevator, 6mm wide, L 200mm	Stainless Steel CE marking	Each
292	Periosteal Elevator, straight shaft, 14mm wide, L 200mm	Stainless Steel CE marking	Each
293	Bone Chisel, 12mm wide	Stainless Steel CE marking	Each
294	Bone Chisel, 18mm wide	Stainless Steel CE marking	Each
295	Bone Gauge, curved, for cancellous bone graft harvest., 5mm wide	Stainless Steel CE marking	Each
296	Bone Gauge, curved, for cancellous bone graft harvest., 10mm wide	Stainless Steel CE marking	Each

M Pandey  
27.02.24

Superintendent

Bhima Bhoi Medical College & Hospital,  
Balangir - 767001



297	Bone Nibbler single action, length 6 inch	Stainless Steel CE marking	Each
298	Bone Nibbler double action, Straight, length 9 inch	Stainless Steel CE marking	Each
299	Bone Nibbler double action, Angular, length 9 inch	Stainless Steel CE marking	Each
300	Cancellous Bone Impactor, round, L 140mm	Stainless Steel CE marking	Each
301	Cancellous Bone Impactor, square, L 140mm	Stainless Steel CE marking	Each
302	Cancellous Bone Impactor, Rectangular, L 140mm	Stainless Steel CE marking	Each
303	Curette small	Stainless Steel CE marking	Each
304	Curette large	Stainless Steel CE marking	Each
305	Osteotome large(12mm size) curved	Stainless Steel CE marking	Each
306	Osteotome large(12mm size) straight	Stainless Steel CE marking	Each
307	Osteotome small(6mm size) curved	Stainless Steel CE marking	Each
308	Osteotome small(6mm size) straight	Stainless Steel CE marking	Each
309	Patella reduction clamp 4 prongs	Stainless Steel CE marking	Each
310	Patella reduction clamp 2 prongs	Stainless Steel CE marking	Each
311	Wire Passer, 70mm bending diameter	Stainless Steel CE marking	Each
312	Wire Tightener with handle and two pegs, L 240mm	Stainless Steel CE marking	Each
313	Holding Forceps for Cerclage Wires, L 170mm	Stainless Steel CE marking	Each
314	Wire Bending Pliers, L 155mm	Stainless Steel CE marking	Each
315	Parallel Pliers, flat nosed, L 160mm	Stainless Steel CE marking	Each
316	Wire Cutter, large, L 220mm	Stainless Steel CE marking	Each
317	Wire Cutter, short, L 175mm	Stainless Steel CE marking	Each
318	Jumbo wire cutter	Stainless Steel CE marking	Each
319	Periarticular reduction forceps with jaw opening up to 80mm, ball tips at pincer end with ratchet lock	Stainless Steel CE marking	Each
320	Kirschner Wires bender 1.25 to 2.5mm dia., L 120mm	Stainless Steel CE marking	Each
321	Plate bender for 3.5 plate	Stainless Steel CE marking	Each Set
322	Plate bender for 4.5 plate	Stainless Steel CE marking	Each Set
323	Plate Bending Plier (Roller type) for Plates 2.4 to 4.5, L 230mm	Stainless Steel CE marking	Each
324	Bending Template for DCP® 3.5 and LC-DCP 3.5, L 87mm	Stainless Steel CE marking	Each
325	Plate bending press for 3.5- and 4.5-mm plate	Stainless Steel CE marking	Each
326	Reduction Forceps with points	Stainless Steel CE marking	Each
327	Reduction Forceps with points, wide, ratchet lock	Stainless Steel CE marking	Each
328	Bone Holding Forceps small	Stainless Steel CE marking	Each
329	Bone Holding Forceps medium	Stainless Steel CE marking	Each



330	Bone Holding Forceps large	Stainless Steel CE marking	Each
331	Hoffman Retractor (bone spike), small,	Stainless Steel CE marking	Each
332	Hoffman Retractor (bone spike), medium	Stainless Steel CE marking	Each
333	Hoffman Retractor (bone spike), large	Stainless Steel CE marking	Each
334	Plate holding forceps straight (small)	Stainless Steel CE marking	Each
335	Plate holding forceps straight (medium)	Stainless Steel CE marking	Each
336	Plate holding forceps straight (large)	Stainless Steel CE marking	Each
337	K nail reamer (non-cannulated size 7 to 12)	Stainless Steel CE marking	Each Set
338	K nail reamer (cannulated size 8 to 12)	Stainless Steel CE marking	Each Set
339	K Nail Extractor	Stainless Steel CE marking	Each
340	Interlocking Nail Extractor	Stainless Steel CE marking	Each
341	Universal adaptor for nail extractor size 6-9	Stainless Steel CE marking	Each
342	Universal adaptor for nail extractor size 9-12	Stainless Steel CE marking	Each
343	Universal adaptor for nail extractor size >12	Stainless Steel CE marking	Each
344	Wrench for nail extractor	Stainless Steel CE marking	Each
345	Heavy Mallet for nail extractor (should be possible to mount on nail extractor)	Stainless Steel CE marking	Each
346	Extraction hook 3.7mm for cannulated nail removal (broken nail removal)	Stainless Steel CE marking	Each
347	Universal chuck with T handle	Stainless Steel CE marking	Each

348	<b>X-RAY VIEW BOX (LED)</b>	Should be FDA/ CE/BIS approved product Manufacturer and Supplier should have ISO 13485 certification for quality standards. <b>TECHNICAL CHARACTERISTICS</b> Should be ultra-thin X ray film illuminator using light It should have a thickness of 30 mm . It should be suitable for viewing 14"x17" film. Should have position to insert 8 films in 2 rows. The LED light must have a life span of more	Each



than 1,00,000 hours.
It should have easy insertion & removal of the film
It should have homogeneous illumination more than 95% and maximum intensity of over 10,000 lux
It should have an on-off switch along with digital feather touch dimmer and a button to set the intensity
It should have fully electronic continuous brightness control, with adjustment range of approximately 90%.
It should have directly connectable to power supply without any external adapter.
It should have flicker free high frequency light for reduction of eye strain.
It should have flicker free high frequency light for reduction of eye strain.
It should have external fuses for protection against power surge.
10 step Digital dimmer facility with step up/step down intensity of 500 lux or less.
Should have automatic film sensor
Should have facility to switch on only the section where the film needs to be viewed
Power Supply:-230V, AC, 50Hz. 3 Amps, Single phase

M. Pande  
27/02/24

Superintendent

Bhima Bhoi Medical College & Hospital  
Balangir - 767001